

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAR 23 2010

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130130</u>	
Name <u>JEREMY CONKLING</u> ID# <u>6715</u> Date <u>03.18.10</u>	
A	Agency <u>VALDEZ POLICE DEPARTMENT</u> Phone # <u>835-4560</u>
Instrument Location <u>VPD JAIL</u>	
Alco S/N <u>70729</u> Target Value <u>.081</u> High Pressure <u>1400</u>	
B	Alco Test Value Average <u>.083</u> <u>.083</u> 1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco
Signature <u>[Signature]</u> 271	
(OVER)	

BMB  
3/24/10

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


(Do Not Write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 15<sup>th</sup> day of April, 2010.

Carolyn M. Noland  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130130 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130130

MARCH 18, 2010

OPERATOR'S NAME:

CONKLING/JERZY/WHEELER

OPERATOR'S NUMBER: 6715

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI:

UOC

O.L. #: UOC

DEPT/AGENCY: UOZ1

CRIME/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: .001 ✓

ALCO S/N: 70729 ✓

### — BREATH ANALYSIS —

.001 ADJUSTED FOR 30.71 in  
ALCO TARGET .003 00:50  
BLANK TEST .000 00:51  
INTERNAL STANDARD VERIFIED 00:51  
ALCO TO 30.71 in .003 00:52  
BLANK TEST .000 00:53  
SUBJECT SAMPLE .000 00:53  
BLANK TEST .000 00:54  
ALCO TO 30.71 in .003 00:54  
BLANK TEST .000 00:55

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130130

MARCH 18, 2010

TIME 00:47

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/28/01  
VENTILATORS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
MANIFOLD: 30.71 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJKL  
MNPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~